

Graduated Pipettes, Hirschmann

Graduated pipettes, not graduated to tip, zero point at the top, with dated batch identification, class AS, with declaration of conformity acc. to DIN 12600, main point ring graduation, AR-GLAS®, clear glass, acc. to DIN EN ISO 835, type 1, graduation and markings in durable amber stain. EX calibrated. For 5 ml and up, equipped with cotton wool stopper as a standard.

Vol.	Graduation	Flow time	Tol. ±	Order No.	PU
ml	ml	s	ml		pc/Pack
0.5	0.01	4 – 10	0.006	14 71 11721	12
1	0.01	4 – 10	0.007	14 71 11727	12
2	0.02	4 – 10	0.01	14 71 11735	12
5	0.05	7 – 13	0.03	14 71 11748	12
10	0.1	7 – 13	0.05	14 71 11758	12
25	0.1	11 – 17	0.1	14 71 11766	6

Flow time: waiting time 5 s.



HIRSCHMANN®
webcode  **WW4ZV005**

