

Graduated Pipettes, Hirschmann

Graduated pipettes, graduated to tip, Schellbach line, with dated batch identification, with declaration of conformity acc. to DIN 12600, class AS, main point ring graduation, AR-GLAS®, clear glass, acc. to DIN EN ISO 835, type 3, graduation and markings in durable blue ceramic stain. EX calibrated. For 5 ml and up, equipped with cotton wool stopper as a standard.

Vol.	Graduation	Flow time	Tol. ±	Order No.	PU
ml	ml	s	ml		pc/Pack
0.5	0.01	4 – 10	0.006	14 71 12221	12
1	0.01	4 – 10	0.007	14 71 12227	12
2	0.02	4 – 10	0.01	14 71 12235	12
5	0.05	7 – 13	0.03	14 71 12248	12
10	0.1	7 – 13	0.05	14 71 12258	12
25	0.1	11 – 17	0.1	14 71 12266	6

Flow time: waiting time 5 s.



HIRSCHMANN®
 **WW4ZV006**

